MAYO CLINIC
Mayo Medical Laboratories

TEST DEFINITION

8/30/2010

Code	Name
200027	Thyroglobulin, Tumor Marker, S

ORDER CODE	EFF DATE	TC	TITLE	Checking Normals	Print normals (# coded)	Perform Site *
200027(profile)			Thyroglobulin, Tumor Marker, S			
201001	8/5/2010		Thyroglobulin Ab Scrn, S			NEL

UNIT CODE ALWAYS MESSAGE - [ZHTG]

The testing method is an immunoenzymatic assay manufactured by Beckman Coulter Inc. and performed on the Unicel DXI 800.
Values obtained with different assay methods or kits may be different and cannot be used interchangeably.
Test results cannot be interpreted as absolute evidence for the presence or absence of malignant disease.
Specimens with thyroglobulin concentrations greater than 250,000 ng/mL may give falsely lower results.

Transport temp : Refrig\Frozen OK\Ambient <7 days OK

201001 Thyroglobulin Ab Scrn, S

Units: IU/mL

NO SEX

All Ages : 0.0-3.9 ; <4.0

MALE

All Ages : 0.0-3.9 ; <4.0

FEMALE

All Ages : 0.0-3.9 ; <4.0

- - - - -

200103	8/5/2010		Thyroglobulin, Tumor Marker, S			NEL

UNIT CODE ALWAYS MESSAGE - [ZHTG]

The testing method is an immunoenzymatic assay manufactured by Beckman Coulter Inc. and performed on the Unicel DXI 800.
Values obtained with different assay methods or kits may be different and cannot be used interchangeably.

```
                         Test results cannot be interpreted as absolute evidence for
                         the presence or absence of malignant disease.
                         Specimens with thyroglobulin concentrations greater than
                         250,000 ng/mL may give falsely lower results.
                  Transport temp : Refrig\Frozen OK\Ambient NO


  ORDER      EFF                                                                        Perform
  CODE       DATE   TC     TITLE          Checking Normals              Print normals (# coded)  Site *
 ------      ----   ----   ----------     ----------------              ----------------------   -------
200027  (continued)
                  200103   Thyroglobulin, Tumor Marker, S
                             Units: ng/mL
                                NO SEX
                             No age        :                              ; #N23745N0
                             1D to 15Y     :                              ; #N23745N0
                             16Y to 150Y   : 0-33                         ; #N23745CB
                                MALE
                             No age        :                              ; #N23745N0
                             1D to 15Y     :                              ; #N23745N0
                             16Y to 150Y   : 0-33                         ; #N23745CB
                                FEMALE
                             No age        :                              ; #N23745N0
                             1D to 15Y     :                              ; #N23745N0
                             16Y to 150Y   : 0-33                         ; #N23745CB
                         -  -  -  -  -
- - - - - - - - - - - - - - - - - - - - - - - - - - -

*Performing Site Legend
===============================================================================================
==
NEL     Mayo Medical Laboratories New England                LAB DIRECTOR:   Lynn A. Cheryk, Ph.D.
        160 Dascomb Road
        Andover, MA  01810
===============================================================================================
==

CODE         TEXT
-----        ------------------
N23745CB     <=33
             Athyrotic individuals
             normally have hTg values
             <=2.
--------
N23745N0     <=33 (>=16 y)
             Athyrotic individuals
             normally have hTg values
             <=2.
```

Total of 2 normals codes

*** End of Report ***