MAYO CLINIC
Mayo Medical Laboratories

TEST DEFINITION

8/31/2010

```
MML                                Name/order code cross-reference
Code   Name
----   --------------------
83069  Thyroglobulin, Tumor Marker, S

MML                                   MML Test setup information
  ORDER      EFF                                                                          Perform
  CODE       DATE   TC     TITLE          Checking Normals              Print normals (# coded)  Site *
 ------      ----   ----   ----------     ----------------              -----------------------  -------
83069(profile)             Thyroglobulin, Tumor Marker, S
   84382     8/5/2010      Thyroglobulin Ab Scrn, S                                              SDL
                   UNIT CODE ALWAYS MESSAGE - [ZHTG]
                          The testing method is an immunoenzymatic assay manufactured
                          by Beckman Coulter Inc. and performed on the Unicel DXI
                          800.
                          Values obtained with different assay methods or kits may be
                          different and cannot be used interchangeably.
                          Test results cannot be interpreted as absolute evidence for
                          the presence or absence of malignant disease.
                          Specimens with thyroglobulin concentrations greater than
                          250,000 ng/mL may give falsely lower results.
                   Transport temp : Refrig\Frozen OK\Ambient NO
                   84382    Thyroglobulin Ab Scrn, S
                               Units: IU/mL
                                  NO SEX
                               All Ages : 0.0-3.9                      ; <4.0
                                  MALE
                               All Ages : 0.0-3.9                      ; <4.0
                                  FEMALE
                               All Ages : 0.0-3.9                      ; <4.0
                          -  -  -  -  -
   23745     8/5/2010      Thyroglobulin, Tumor Marker, S                                        SDL
                   UNIT CODE ALWAYS MESSAGE - [ZHTG]
                          The testing method is an immunoenzymatic assay manufactured
                          by Beckman Coulter Inc. and performed on the Unicel DXI
                          800.
                          Values obtained with different assay methods or kits may be
                          different and cannot be used interchangeably.
                          Test results cannot be interpreted as absolute evidence for
                          the presence or absence of malignant disease.
                          Specimens with thyroglobulin concentrations greater than
```

250,000 ng/mL may give falsely lower results.
Transport temp : Refrig\Frozen OK\Ambient NO

```
MML                                  MML Test setup information
  ORDER      EFF                                                                          Perform
  CODE       DATE   TC     TITLE          Checking Normals              Print normals (# coded)  Site *
 ------      ----   ----   ----------     ----------------              -----------------------  -------
83069  (continued)
                    23745    Thyroglobulin, Tumor Marker, S
                                Units: ng/mL
                                   NO SEX
                                No age        :                          ; #N23745N0
                                1D to 15Y     :                          ; #N23745N0
                                16Y to 150Y   : 0-33                     ; #N23745CB
                                   MALE
                                No age        :                          ; #N23745N0
                                1D to 15Y     :                          ; #N23745N0
                                16Y to 150Y   : 0-33                     ; #N23745CB
                                   FEMALE
                                No age        :                          ; #N23745N0
                                1D to 15Y     :                          ; #N23745N0
                                16Y to 150Y   : 0-33                     ; #N23745CB
                            -  -  -  -  -
- - - - - - - - - - - - - - - - - - - - - - - - - - - - - -

*Performing Site Legend
==========================================================================================================
SDL     Mayo Clinic Dpt of Lab Med & Pathology              LAB DIRECTOR:   Franklin R. Cockerill, III, M.D.
        Superior Dr.
        3050 Superior Dr. NW
        Rochester, MN  55901
==========================================================================================================

MML                                                   Messages used as normals
CODE         TEXT
-----        ------------------
N23745CB     <=33
             Athyrotic individuals
             normally have hTg values
             <=2.
--------
N23745N0     <=33 (>=16 y)
             Athyrotic individuals
             normally have hTg values
             <=2.
--------

Total of 2 normals codes

    *** End of Report ***
```