MAYO CLINIC
Mayo Medical Laboratories

TEST DEFINITION

9/29/2010

```
MML                                Name/order code cross-reference
Code   Name
----   --------------------
57193  Limulus Amebocyte Lysate Assay, QN

MML                                   MML Test setup information
  ORDER      EFF                                                                                  Perform
  CODE       DATE   TC     TITLE          Checking Normals              Print normals (# coded)  Site *
 ------      ----   ----   ----------     ----------------              -----------------------  -------
57193(profile)             Limulus Amebocyte Lysate Assay, QN
   80772     9/22/2010     Limulus Lysate Assay
                   UNIT CODE ALWAYS MESSAGE - [Z80772N1]
                          LEVEL DETECTED              INTERPRETATION
                          ------------------------------------------------------------
                          <0.05 EU/mL                 None Detected (Negative)
                           0.25 EU/mL                 USP acceptable limits for
                                                      injectable or irrigation water.
                           0.50 EU/mL                 USP acceptable limits for
                                                      inhalatory water.
                           2.00 EU/mL                 Acceptable upper limit for
                                                      Hemodialysis reuse water.

                          The LAL is used as a quantitative test to detect gram-
                          negative endotoxin in aqueous solutions used in patient
                          management. The LAL assay is not recommended for serum or
                          plasma samples due to the presence of inhibitory factors. It
                          is essential to maintain specimen sterility and prevent
                          false positive results from exogenous gram negative
                          bacteria.

                          Test performed by: Focus Diagnostics, Inc.
                                             5785 Corporate Avenue
                                             Cypress, CA 90630-4750
                   Transport temp : Frozen\Refrig NO\Ambient NO
                   90922    Specimen Type
                          -  -  -  -  -
```

ORDER CODE	EFF DATE	TC	TITLE	Checking Normals	Print normals (# coded)	Perform Site *

57193 (continued)

90923 Diluent

- - - - -

90924 Pharmacopeia Endotoxin Limit

- - - - -

25290 Limulus Lysate

Units: EU/mL

- - - - -

57190 9/22/2010 Limulus Lysate

UNIT CODE ALWAYS MESSAGE - [Z80772N1]

LEVEL DETECTED	INTERPRETATION
<0.05 EU/mL	None Detected (Negative)
0.25 EU/mL	USP acceptable limits for injectable or irrigation water.
0.50 EU/mL	USP acceptable limits for inhalatory water.
2.00 EU/mL	Acceptable upper limit for Hemodialysis reuse water.

The LAL is used as a quantitative test to detect gram-negative endotoxin in aqueous solutions used in patient management. The LAL assay is not recommended for serum or plasma samples due to the presence of inhibitory factors. It is essential to maintain specimen sterility and prevent false positive results from exogenous gram negative bacteria.

Test performed by: Focus Diagnostics, Inc.
5785 Corporate Avenue
Cypress, CA 90630-4750

Transport temp : Frozen\Refrig NO\Ambient NO

22172 Limulus Lysate

Units: EU/mL

- - - - -

57191 9/22/2010 Limulus Lysate

UNIT CODE ALWAYS MESSAGE - [Z80772N1]

LEVEL DETECTED	INTERPRETATION
<0.05 EU/mL	None Detected (Negative)
0.25 EU/mL	USP acceptable limits for injectable or irrigation water.
0.50 EU/mL	USP acceptable limits for inhalatory water.
2.00 EU/mL	Acceptable upper limit for Hemodialysis reuse water.

The LAL is used as a quantitative test to detect gram-
negative endotoxin in aqueous solutions used in patient
management. The LAL assay is not recommended for serum or
plasma samples due to the presence of inhibitory factors. It
is essential to maintain specimen sterility and prevent
false positive results from exogenous gram negative
bacteria.

Test performed by: Focus Diagnostics, Inc.
5785 Corporate Avenue
Cypress, CA 90630-4750
Transport temp : Frozen\Refrig NO\Ambient NO

MML MML Test setup information

ORDER CODE	EFF DATE	TC	TITLE	Checking Normals	Print normals (# coded)	Perform Site *
57193 (continued)						
		22173	Limulus Lysate Units: EU/mL			
			- - - - -			
57192	9/22/2010		Limulus Lysate			

UNIT CODE ALWAYS MESSAGE - [Z80772N1]

LEVEL DETECTED	INTERPRETATION
<0.05 EU/mL	None Detected (Negative)
0.25 EU/mL	USP acceptable limits for injectable or irrigation water.
0.50 EU/mL	USP acceptable limits for inhalatory water.
2.00 EU/mL	Acceptable upper limit for Hemodialysis reuse water.

The LAL is used as a quantitative test to detect gram-
negative endotoxin in aqueous solutions used in patient
management. The LAL assay is not recommended for serum or
plasma samples due to the presence of inhibitory factors. It
is essential to maintain specimen sterility and prevent
false positive results from exogenous gram negative
bacteria.

Test performed by: Focus Diagnostics, Inc.
5785 Corporate Avenue
Cypress, CA 90630-4750
Transport temp : Frozen\Refrig NO\Ambient NO
22174 Limulus Lysate
Units: EU/mL
- - - - -
- -

MML Messages used as normals

CODE	TEXT

Total of 0 normals codes

*** End of Report ***