



MAYO CLINIC
 Mayo Medical Laboratories

TEST DEFINITION

12/23/2010

Code Name

 9290 D-Dimer, P

**** End of Name to order code Report ****

MML MML Test setup information

ORDER CODE	EFF DATE	TC	TITLE	Checking Normals	Print normals (# coded)	Perform Site *
9290	11/9/2010		D-Dimer, P			MCR
			Transport temp : Frozen\Refrig NO\Ambient NO			
			28579 Fibrinogen Equivalent Units (FEU)			
			Units: mcg/mL FEU			
			NO SEX			
			All Ages : 0-0.50		; <=0.50	
			MALE			
			All Ages : 0-0.50		; <=0.50	
			FEMALE			
			All Ages : 0-0.50		; <=0.50	

		9290	D-Dimer Units (DDU)			
			Units: ng/mL DDU			
			NO SEX			
			All Ages : 0-250		; #N9290N	
			MALE			
			All Ages : 0-250		; #N9290N	
			FEMALE			
			All Ages : 0-250		; #N9290N	
			TEST CODE ALWAYS MESSAGE - [N9290N]			
			N9290N <=250			
			The stated reference value is the cutoff to use D-Dimer for exclusion of deep vein thrombosis and/or pulmonary embolism.			

MML

MML Test setup information

*Performing Site Legend

=====
==
MCR Mayo Clinic Dpt of Lab Med & Pathology LAB DIRECTOR: Franklin R. Cockerill, III, M.D.
 200 First Street SW
 Rochester, MN 55905
=====

*** End of Report ***

MML

Messages used as normals

CODE	TEXT
N9290N	<=250 The stated reference value is the cutoff to use D-Dimer for exclusion of deep vein thrombosis and/or pulmonary embolism.

Total of 1 normals codes

*** End of Report ***