



TEST DEFINITION

12/28/2010

MML Name/order code cross-reference

Code Name

81130 HCV RNA Quantification, S

MML MML Test setup information

ORDER CODE	EFF DATE	TC	TITLE	Checking Normals	Print normals (# coded)	Perform Site *
81130	11/17/2010		HCV RNA Quantification, S			SDL
			Transport temp : Frozen\Refrig NO\Ambient NO			
			81130 HCV RNA Quantification, S			
			Units: IU/mL			
			NO SEX			
			All Ages :		; <615	
			MALE			
			All Ages :		; <615	
			FEMALE			
			All Ages :		; <615	
			TEST CODE ALWAYS MESSAGE - [Z81130A]			
			Z81130A Quantification range of this assay is 615 to 7,690,000 IU/mL			
			(2.79 to 6.89 log IU/mL).			
			Testing was done by the VERSANT HCV RNA 3.0 Assay (bDNA)			
			(Siemens Healthcare Diagnostics Inc.).			

*Performing Site Legend

```

=====
SDL      Mayo Clinic Dpt of Lab Med & Pathology      LAB DIRECTOR:  Franklin R. Cockerill, III, M.D.
          Superior Dr.
          3050 Superior Dr. NW
          Rochester, MN 55901
=====

```

MML Messages used as normals

CODE TEXT

Total of 0 normals codes

*** End of Report ***