MAYO CLINIC
Mayo Medical Laboratories

TEST DEFINITION

1/28/2011

Code	Name
800066	PSA Total and Free, S

**** End of Name to order code Report ****

MML MML Test setup information

ORDER CODE	EFF DATE	TC	TITLE	Checking Normals	Print normals (# coded)	Perform Site *
800066	1/11/2011		PSA Total and Free, S			MCF

UNIT CODE ALWAYS MESSAGE - [ZPSAA]

The testing method is an electrochemiluminescence assay manufactured by Roche Diagnostics Inc. and performed on the Modular or Cobas system. Values obtained with different assay methods or kits may be different and cannot be used interchangeably.

Test results cannot be interpreted as absolute evidence for the presence or absence of malignant disease.

Transport temp : Frozen OK\Refrig NO\Ambient NO

800066 Total PSA

Units: ng/mL

NO SEX

No age	:	; #N92840Y
1D to 39Y	:	; #N9284A39
40Y to 49Y	:	; #N9284A49
50Y to 59Y	:	; #N9284A59
60Y to 69Y	:	; #N9284A69
70Y to 79Y	:	; #N9284A79
80Y to 150Y	:	; #N9284A150

MALE

No age	:	; #N92840YM
1D to 39Y	: 0.0-2.0	; <=2.0
40Y to 49Y	: 0.0-2.5	; <=2.5
50Y to 59Y	: 0.0-3.5	; <=3.5
60Y to 69Y	: 0.0-4.5	; <=4.5
70Y to 79Y	: 0.0-6.5	; <=6.5

(continued next page)

```
  ORDER      EFF                                                                          Perform
  CODE       DATE   TC     TITLE          Checking Normals              Print normals (# coded)  Site *
 ------      ----   ----   ----------     ----------------              -----------------------  -------
800066  (continued)
                    800066   Total PSA  (continued)
                                80Y to 150Y   : 0.0-7.2                  ; <=7.2
                                   FEMALE
                                All Ages :                               ; Not applicable
                           -  -  -  -  -
                    800067   Free PSA
                                Units: ng/mL
                           -  -  -  -  -
                    800068   Free PSA/PSA Ratio
                           -  -  -  -  -
- - - - - - - - - - - - - - - - - - - - - - - - - - - - -
```

*Performing Site Legend

MCF Mayo Clinic Jacksonville Clinical Lab LAB DIRECTOR: Arthur D. Jones, Jr. M.D.
4500 San Pablo Rd
Jacksonville, FL 32224

*** End of Report ***


```
CODE         TEXT
-----        ------------------
N92840Y      Males:
             <=2.0 (<40 y)
             <=2.5 (40-49 y)
             <=3.5 (50-59 y)
             <=4.5 (60-69 y)
             <=6.5 (70-79 y)
             <=7.2 (>=80 y)

             Females:
             Not applicable
--------
N92840YM     <=2.0 (<40 y)
             <=2.5 (40-49 y)
             <=3.5 (50-59 y)
             <=4.5 (60-69 y)
```

```
             <=6.5 (70-79 y)
             <=7.2 (>=80 y)
--------
N9284A150    Males:
             <=7.2 (>=80 y)

             Females:
             Not applicable
--------
N9284A39     Males:
             <=2.0 (<40 y)

             Females:
             Not applicable
--------
N9284A49     Males:
             <=2.5 (40-49 y)

             Females:
             Not applicable
--------
N9284A59     Males:
             <=3.5 (50-59 y)

             Females:
             Not applicable
--------
N9284A69     Males:
             <=4.5 (60-69 y)

             Females:
             Not applicable
--------
N9284A79     Males:
             <=6.5 (70-79 y)

             Females:
             Not applicable
--------


Total of 8 normals codes

    *** End of Report ***
```