MAYO CLINIC
Mayo Medical Laboratories

TEST DEFINITION

10/13/2011

ORDER CODE	EFF DATE	TC	TITLE	CHECKING NORMALS	PRINT NORMALS (# CODED)	PERFORM SITE *
800126(PROFILE)			THYROGLOBULIN, TUMOR MARKER, S			
800312	10/6/2011		THYROGLOBULIN ANTIBODY SCREEN, S			MCF

UNIT CODE ALWAYS MESSAGE - [ZHTG2]

THE THYROGLOBULIN TESTING METHOD IS AN IMMUNOENZYMATIC ASSAY MANUFACTURED BY BECKMAN COULTER INC. AND PERFORMED ON THE UNICEL DXI 800.
THE THYROGLOBULIN ANTIBODY TESTING METHOD IS AN ELECTROCHEMILUMINESCENCE ASSAY MANUFACTURED BY ROCHE DIAGNOSTICS INC. AND PERFORMED ON THE MODULAR OR COBAS SYSTEM.
VALUES OBTAINED FROM DIFFERENT ASSAY METHODS OR KITS MAY BE DIFFERENT AND CANNOT BE USED INTERCHANGEABLY.

THE RESULTS CANNOT BE INTERPRETED AS ABSOLUTE EVIDENCE FOR THE PRESENCE OR ABSENCE OF MALIGNANT DISEASE.

SPECIMENS WITH THYROGLOBULIN CONCENTRATIONS GREATER THAN 250,000 NG/ML MAY GIVE FALSELY LOWER RESULTS.

TRANSPORT TEMP : REFRIG\FROZEN OK\AMBIENT NO

800312 THYROGLOBULIN ANTIBODY SCREEN, S

UNITS: IU/ML

NO SEX
ALL AGES : 0-21 ; <22

MALE
ALL AGES : 0-21 ; <22

FEMALE
ALL AGES : 0-21 ; <22

- - - - -

(CONTINUED NEXT PAGE)

ORDER CODE	EFF DATE	TC	TITLE	CHECKING NORMALS	PRINT NORMALS (# CODED)	PERFORM SITE *
800126 (CONTINUED)						
800131	10/6/2011		THYROGLOBULIN, TUMOR MARKER, S			MCF

UNIT CODE ALWAYS MESSAGE - [ZHTG2]

THE THYROGLOBULIN TESTING METHOD IS AN IMMUNOENZYMATIC ASSAY MANUFACTURED BY BECKMAN COULTER INC. AND

performed on the Unicel DXI 800.
The thyroglobulin antibody testing method is an
electrochemiluminescence assay manufactured by Roche
Diagnostics Inc. and performed on the Modular or Cobas
System.
Values obtained from different assay methods or kits
may be different and cannot be used interchangeably.

The results cannot be interpreted as absolute evidence
for the presence or absence of malignant disease.

Specimens with thyroglobulin concentrations greater
than 250,000 ng/mL may give falsely lower results.

Transport temp : Refrig\Frozen OK\Ambient NO

800131 Thyroglobulin, Tumor Marker, S

Units: ng/mL

NO SEX

No age	:		; #N23745N0
1D to 15Y	:		; #N23745N0
16Y to 150Y	:	0-33	; #N23745CB

MALE

No age	:		; #N23745N0
1D to 15Y	:		; #N23745N0
16Y to 150Y	:	0-33	; #N23745CB

FEMALE

No age	:		; #N23745N0
1D to 15Y	:		; #N23745N0
16Y to 150Y	:	0-33	; #N23745CB

- - - - -

- -

*Performing Site Legend

==========

MCF Mayo Clinic Jacksonville Clinical Lab LAB DIRECTOR: Gretchen Johns, M.D.
4500 San Pablo Rd
Jacksonville, FL 32224

==========

*** End of Report ***

CODE	TEXT
N23745CB	<=33 Athyrotic individuals normally have hTg values <=2.
N23745N0	<=33 (>=16 y) Athyrotic individuals

NORMALLY HAVE HTG VALUES
<=2.

TOTAL OF 2 NORMALS CODES

*** END OF REPORT ***