

Patient Name TEST,IMPLEMENTATION TESTING	Patient ID 321	Age 61	Gender M	Order # M101351772
Ordering Phys TESTING				DOB 12/07/1950
Client Order # M101351772	Account Information			Report Notes
Collected 02/02/2012 12:00	C7028846-DLMP ROCHESTER 3050 SUPERIOR DRIVE ROCHESTER,MN 55901			
Printed 02/02/2012 14:43				

Test	Flag	Results	Unit	Reference Value	Perform Site*
CMV Quant PCR		<250	REPORTED 02/02/2012 13:56 copies/mL		Y10 4

<250 copies/mL - Negative or Not Detected
 250 - 500 copies/mL - Qualitative positive but below limit of quantitation.
 501 - 1,000,000 copies/mL - Quantitative positive
 >1,000,000 copies/mL - Greater than maximum quantitative range.

REFERENCE NOTE: This test employs real-time amplification of a CMV-specific conserved genetic target. A positive result should be coupled with clinical indicators for diagnosis. A negative CMVPCR result (<250 copies/mL) does not exclude CMV involvement in a disease process. This test was developed and its performance characteristics determined by the UCH Clinical Laboratory. It has not been cleared or approved by U.S. Food and Drug Administration. The FDA has determined that such clearance or approval is not necessary. This test is used for clinical purposes. It should not be regarded as investigational or for research. This laboratory is certified under the Clinical Laboratory Improvement Amendment of 1988 ("CLIA") as qualified to perform high complexity clinical testing.

* Performing Site:

Y104	University of Colorado Hospital Clinical Laboratory 12401 E 17th Ave Aurora, CO 80045	Lab Director:
------	--	---------------

Patient Name TEST,IMPLEMENTATION TESTING	Collection Date and Time 02/02/2012 12:00	Report Status Final
Page 1 of 1		** End of Report **

* Report times for Mayo performed tests are CST/CDT