

Patient Name SAMPLEREP,CEAPT ABN	Patient ID SA00048245	Age 45	Gender F	Order # SA00048245
Ordering Phys				DOB 06/10/1966
Client Order # SA00048245	Account Information			Report Notes
Collected 08/06/2011 13:00	C7028846-DLMP ROCHESTER 3050 SUPERIOR DRIVE ROCHESTER,MN 55901			
Printed 08/27/2012 13:45				

Test	Flag	Results	Unit	Reference Value	Perform Site*
CEA, Peritoneal Fluid		8.0	ng/mL	REPORTED 08/07/2012 08:15	SDL
<p>A peritoneal fluid CEA concentration > 6.0 ng/mL is suspicious but not diagnostic of a malignant source of the ascites. This cutoff yielded 48% sensitivity, 99% specificity in a study of 137 patients presenting with ascites. CEA levels in peritoneal fluid are most informative when interpreted in conjunction with the CEA serum levels. Tumor marker tests are not specific for malignancy. This test result should be interpreted in the context of clinical presentation, imaging and cytology findings.</p> <p>The testing method is an immunoenzymatic assay manufactured by Beckman Coulter Inc. and performed on the UniCel DxI 800.</p> <p>Values obtained with different assay methods or kits may be different and cannot be used interchangeably.</p> <p>Test results cannot be interpreted as absolute evidence for the presence or absence of malignant disease.</p>					
Site		Right Side			SDL

* Performing Site:

SDL	Mayo Clinic Laboratories - Rochester Superior Drive 3050 Superior Dr. NW Rochester, MN 55901	Lab Director:
-----	---	---------------

Patient Name SAMPLEREP,CEAPT ABN	Collection Date and Time 08/06/2011 13:00	Report Status Final
Page 1 of 1		** End of Report **

* Report times for Mayo performed tests are CST/CDT