

Patient Name TEST,IMPLEMENTATION TESTING	Patient ID 321	Age 38	Gender M	Order # R1057038
Ordering Phys unknown				DOB 10/14/1974
Client Order # R1057038	Account Information C7028845-DLMP Jacksonville SDSC 2 - Client Support Rochester, MN 55901			Report Notes
Collected 03/01/2013 11:00				
Printed 03/27/2013 08:02				

Test	Flag	Results	Unit	Reference Value	Perform Site*
Thyroglobulin, Tumor Marker, S				REPORTED 03/01/2013 12:49	
Thyroglobulin, Tumor Marker, S		32.9	ng/mL		MCF
-- REFERENCE VALUE --					
<=33					
Athyrotic individuals normally have hTg values <=2.					
Thyroglobulin Antibody Screen, S		21	IU/mL	<22	MCF
The thyroglobulin testing method is an immunoenzymatic assay manufactured by Beckman Coulter Inc. and performed on the Unicel DXI 800.					

The thyroglobulin antibody testing method is an electrochemiluminescence assay manufactured by Roche Diagnostics Inc. and performed on the Modular or Cobas System.

Values obtained from different assay methods or kits may be different and cannot be used interchangeably.

The results cannot be interpreted as absolute evidence for the presence or absence of malignant disease.

Specimens with thyroglobulin concentrations greater than 250,000 ng/mL may give falsely lower results.

* Performing Site:

MCF	Mayo Clinic Jacksonville Clinical Lab 4500 San Pablo Road Jacksonville, FL 32224	Lab Director: Gretchen Johns, M.D.
-----	---	------------------------------------

Patient Name TEST,IMPLEMENTATION TESTING	Collection Date and Time 03/01/2013 11:00	Report Status Final
Page 1 of 1		** End of Report **

* Report times for Mayo performed tests are CST/CDT