MAYO CLINIC
Mayo Medical Laboratories

Laboratory Service Report 1-800-533-1710

Patient Name	Patient ID	Age	Gender	Order #
SAMPLEREPORTS,CMVQU POSITIVE	NA00025893	23	M	NA00025893
Ordering Phys CLIENT,CLIENT				**DOB** 05/05/1990
Client Order # NA00025893	**Account Information**			**Report Notes**
Collected 05/13/2013 14:37	C7028847-DLMP New England SDSC 2 - Client Support Rochester, MN 55901			
Printed 05/22/2013 12:43				

Test	Flag	Results	Unit	Reference Value	Perform Site*
			REPORTED 05/13/2013 13:55		
CMV DNA Detect/Quant, P	AB	5000	IU/mL	Undetected	NEL

Result in log IU/mL is 3.70.
The quantification range of this assay is 137 to 9,100,000 IU/mL (2.14 log to 6.96 log IU/mL) with a limit of detection at 91 IU/mL (1.96 log IU/mL). Testing was performed by the COBAS AmpliPrep/COBAS TaqMan CMV Test (Roche Molecular Systems, Inc.). To allow re-baseline with the new test method and comparison of test results until 8-1-2013, the previous laboratory-developed assay (Cytomegalovirus PCR, Quant, P) will be performed at no charge and result will follow.

			REPORTED 05/13/2013 14:01		
Cytomegalovirus PCR, Quant, P	AB	Detected		None detected	NEL
Copies/mL		12000	copies/mL		NEL
Result in log copies/mL = 4.08					
Lower 95% Confidence Interval		5238	copies/mL		NEL
Upper 95% Confidence Interval		27490	copies/mL		NEL

Analyte Specific Reagent: This test was developed and its performance characteristics determined by Mayo Clinic. It has not been cleared or approved by the U.S. Food and Drug Administration.

* Performing Site:

NEL	Mayo Medical Laboratories New England 160 Dascomb Road Andover, MA 01810	Lab Director: Lynn A. Cheryk, Ph.D.

Patient Name	Collection Date and Time	Report Status
SAMPLEREPORTS,CMVQU POSITIVE	05/13/2013 14:37	Final
Page 1 of 1		** End of Report **

* Report times for Mayo performed tests are CST/CDT