

Patient Name TEST,IMPLEMENTATION TESTING	Patient ID 321	Age 1M	Gender F	Order # R1058475
Ordering Phys UNKNOWN,PROVIDER				DOB 10/09/2013
Client Order # R1058475	Account Information			Report Notes
Collected 11/13/2013 06:30	C7028845-DLMP Jacksonville SDSC 2 - Client Support			
Printed 11/13/2013 13:35	Rochester, MN 55901			

Test	Flag	Results	Unit	Reference Value	Perform Site*
CMV DNA Detect/Quant, P		>9100000	IU/mL	Undetected	MCF
RECEIVED: 11/13/2013 07:10 REPORTED: 11/13/2013 09:26 Result in log IU/mL is >6.96. CMV DNA level is >9,100,000 IU/mL (6.96 log IU/mL). This assay cannot accurately quantify CMV DNA above this level. The quantification range of this assay is 137 to 9,100,000 IU/mL (2.14 log to 6.96 log IU/mL) with a limit of detection at 91 IU/mL (1.96 log IU/mL). Testing was performed by the COBAS Ampliprep/COBAS TaqMan CMV Test (Roche Molecular Systems, Inc.).					

* Performing Site:

MCF	Mayo Clinic Jacksonville Clinical Lab 4500 San Pablo Road Jacksonville, FL 32224	Lab Director: Gretchen Johns, M.D.
-----	-------------------------------------------------------------------------------------	------------------------------------

Patient Name TEST,IMPLEMENTATION TESTING	Collection Date and Time 11/13/2013 06:30	Report Status Final
Page 1 of 1		** End of Report **

* Report times for Mayo performed tests are CST/CDT