MAYO
Mayo Medical Laboratories
1-800-533-1710

NEW TEST ANNOUNCEMENT
NOTIFICATION DATE: March 28, 2014
EFFECTIVE DATE: May 6, 2014

Cryptococcus Antigen Titer, LFA, Serum
Test ID: SLFAT

EXPLANATION: *Cryptococcus* Antigen Titer by lateral flow assay (SLFAT) will be replacing the current *Cryptococcus* Ag (SCRYR) by latex agglutination due to increased sensitivity demonstrated by the lateral flow methodology.

USEFUL FOR: Monitoring *Cryptococcus* antigen titers in serum aiding in the diagnosis of cryptococcosis.

METHODOLOGY: Lateral Flow Assay (LFA)
NOTE: End point titers are not interchangeable and do not correlate between the latex agglutination (LA) and lateral flow methods (LFA). Providers should transition to monitor end-point titers by the LFA method. End-point titer values by the LA method will be reported alongside the LFA titer result for comparison until August 2014, at no charge. After this date, only the LFA titer will be performed.

REFERENCE VALUES: Negative

SPECIMEN REQUIREMENTS:
Container/Tube:
Preferred: Red top
Acceptable: Serum gel
Specimen Volume: 0.5 mL
Minimum Volume: 0.3 mL

SPECIMEN STABILITY INFORMATION:

Specimen Type	Temperature	Time
Serum	Refrigerated (preferred)	14 days
	Frozen	14 days

CPT CODE: 87899

DAY(S) SET UP: Monday through Sunday. **ANALYTIC TIME**: 1 day.

CAUTIONS:
- *Cryptococcus* antigen titers acquired by the LFA may be higher than titers achieved by other *Cryptococcus* antigen assays. Titers acquired by different assay methods are not interchangeable.
- *Cryptococcus* antigen titers should be followed using the same assay.
- A positive result is indicative of cryptococcosis, however all test results should be reviewed in light of other clinical findings.
- Testing should not be performed as a screening procedure for the general populations and should only be performed when clinical evidence suggests the diagnosis of cryptococcal disease.

- Testing hemolyzed serum samples may lead to false negative results due to the high background color on the LFA strip.
- Although rare, extremely high concentrations of cryptococcal antigen can result in weak test lines and in extreme instances, yield negative test results.
- This assay has not been evaluated for cross-reactivity in patients with trichosporonosis.

QUESTIONS: Contact your Mayo Medical Laboratories' Regional Manager or
Dunisha Messmer, MML Laboratory Technologist Resource Coordinator
Telephone: 800-533-1710